

Florida Department of Law Enforcement Alcohol Testing Program

CERTIFICATE OF ASSURANCE

The Florida Department of Law Enforcement Alcohol Testing Program approves alcohol reference solution lots used to test the accuracy and precision of evidentiary breath test instruments. Alcohol reference solution lots are prepared by the Department or by a source approved by the Department, and their analysis is important to establish the reliability of alcohol breath test results.

Lot Number: 200403F
Simulated Breath Alcohol Concentration (g/210L): 0.08 g/210L
Date of Approval: April 11, 2004

Source: Alcohol Countermeasure Systems, Inc.
Date of Manufacture: March 26, 2004
For Use Through: March 26, 2006
Number of Bottles Manufactured: 1687

This certifies that ten sample bottles from the above lot were subjected to duplicate gas chromatographic analyses of their alcohol concentrations, that all the results were within the acceptable range, and that this lot of alcohol reference solution is in compliance with the requirements of Chapter 11D-8, F.A.C. and is approved for distribution and use in the State of Florida.

Laura J. Barfield
Signature

Laura J. Barfield
Name (Printed or Typed)

Program Manager
Title

4-13-2004
Date