

Florida Department of Law Enforcement Alcohol Testing Program

CERTIFICATE OF ASSURANCE (Supplemental)

The Florida Department of Law Enforcement Alcohol Testing Program approves alcohol reference solution lots used to test the accuracy and precision of evidentiary breath test instruments. Alcohol reference solution lots are prepared by the Department or by a source approved by the Department, and their analysis is important to establish the reliability of alcohol breath test results.

Lot Number: 200411D
Simulated Breath Alcohol Concentration (g/210L): 0.20 g/210L
Date of Approval: November 28, 2004

Source: Alcohol Countermeasure Systems, Inc.
Date of Manufacture: November 4, 2004
For Use Through: November 4, 2006
Number of Bottles Manufactured: 1679

This certifies that ten sample bottles from the above lot were subjected to duplicate gas chromatographic analyses of their alcohol concentrations, that all the results were within the acceptable range, and that this lot of alcohol reference solution is in compliance with the requirements of Chapter 11D-8, F.A.C. and is approved for distribution and use in the State of Florida.

Laura D Barfield
Signature
Laura D Barfield
Name (Printed or Typed)

Program Manager
Title
12-27-2004
Date