Florida Department of Law Enforcement
Alcohol Testing Program

CERTIFICATE OF ASSURANCE

The Florida Department of Law Enforcement Alcohol Testing Program approves alcohol reference solution lots used to test the accuracy and precision of evidentiary breath test instruments. Alcohol reference solution lots are prepared by the Department or by a source approved by the Department, and their analysis is important to establish the reliability of alcohol breath test results.

Lot Number: 200509A
Simulated Breath Alcohol Concentration (g/210L): 0.05 g/210L
Date of Approval: November 17, 2005

Source: Alcohol Countermeasure Systems, Inc.
Date of Manufacture: September 22, 2005
For Use Through: September 22, 2007
Number of Bottles Manufactured: 1,666

This certifies that ten sample bottles from the above lot were subjected to duplicate gas chromatographic analyses of their alcohol concentrations, that all the results were within the acceptable range, and that this lot of alcohol reference solution is in compliance with the requirements of Chapter 11D-8, F.A.C. and is approved for distribution and use in the State of Florida.

Laura D Barfield — Signature
Program manager — Title
Laura D. Barfield — Name (Printed or Typed)
11-17-05 — Date

FDLE/ATP Form 32 – Revised March, 2001