

SN# 80-000872



Scott Specialty Gases, Inc.

CERTIFICATE OF ANALYSIS EBS™ - Ethanol Breath Standard

Customer Name CMI 316 East Ninth Street Owensboro, KY 42303	Assay Laboratory Scott Specialty Gases 6141 Easton Road, Bldg. 1 Plumsteadville, Pa. 18949	Purchase Order: 368474 Scott Project #:0130363 Lot No.: 5126011 Expiration Date: May 10, 2007
--	---	---

REFERENCE STANDARD	CYLINDER #	CONCENTRATION ppm
N.I.S.T. Traceable Standards*	ALM-060154	103.6
	ALM-060204	118.1
	ALM-059962	222.2
	ALM-037472	251.0

CONTAINER TYPE:105

LOT CERTIFIED VALUE CONCENTRATION

COMPONENTS	ppm (BrAC)	ACCURACY
Ethanol	208.4 (0.080)	± 2%
Nitrogen	Balance	

*Certification: Traceable to N.I.S.T. RGM Ethanol Standards

Floyd Bandy

Analyst:

Date: 05/10/2005