



Scott Specialty Gases, Inc.

CERTIFICATE OF ANALYSIS EBS™ - Ethanol Breath Standard

Customer Name
CMI
316 East Ninth Street
Owensboro, KY 42303

Assay Laboratory
Scott Specialty Gases
6141 Easton Road, Bldg. 1
Plumsteadville, Pa. 18949

Purchase Order: 368833
Scott Project #: 0135169
Lot No.: 520201I
Expiration Date:
July 26, 2007

REFERENCE STANDARD

N.I.S.T. Traceable Standards*

CYLINDER

ALM-060154
ALM-060204
ALM-059962
ALM-037472

CONCENTRATION ppm

103.6
118.1
222.2
251.0

CONTAINER TYPE:105

LOT CERTIFIED VALUE CONCENTRATION

COMPONENTS

Ethanol
Nitrogen

ppm (BrAC)

208.4 (0.080)
Balance

ACCURACY

+/- 2%

*Certification: Traceable to N.I.S.T. RGM Ethanol Standards

Analyst:

Floyd Bards

Date: 07/26/2005