

Installed: 06/22/07
Index - SWP 80700873



Scott Specialty Gases, Inc.

CERTIFICATE OF ANALYSIS EBS™ - Ethanol Breath Standard

Customer Name
CMI
316 East Ninth Street
Owensboro, KY 42303

Assay Laboratory
Scott Specialty Gases
6141 Easton Road, Bldg. 1
Plumsteadville, Pa. 18949

Purchase Order: 369109
Scott Project #: 0138832
Lot No.: 5252021
Expiration Date:
September 14, 2007

REFERENCE STANDARD
N.I.S.T. Traceable Standards*

CYLINDER #
ALM-060154
ALM-060204
ALM-059962
ALM-037472

CONCENTRATION ppm
103.6
118.1
222.2
251.0

CONTAINER TYPE:105

LOT CERTIFIED VALUE
CONCENTRATION

COMPONENTS	ppm (BrAC)	ACCURACY
Ethanol	208.4 (0.080)	+/- 2%
Nitrogen	Balance	

*Certification: Traceable to N.I.S.T. RGM Ethanol Standards

Analyst:

Date: 10/10/2005