



# Scott Specialty Gases, Inc.

## CERTIFICATE OF ANALYSIS

### EBS™ - Ethanol Breath Standard

Customer Name  
CMI  
316 East Ninth Street  
Owensboro, KY 42303

Assay Laboratory  
Scott Specialty Gases  
6141 Easton Road, Bldg. 1  
Plumsteadville, Pa. 18949

Purchase Order: 369020  
Scott Project #: 0139982  
Lot No.: 531401G  
Expiration Date:  
November 14, 2007

<u>REFERENCE STANDARD</u>	<u>CYLINDER #</u>	<u>CONCENTRATION ppm</u>
N.I.S.T. Traceable Standards*	ALM-060154	103.6
	ALM-060204	118.1
	ALM-059962	222.2
	ALM-037472	251.0

CONTAINER TYPE:105

#### LOT CERTIFIED VALUE CONCENTRATION

<u>COMPONENTS</u>	<u>ppm (BrAC)</u>	<u>ACCURACY</u>
Ethanol	208.4 (0.080)	+/- 2%
Nitrogen	Balance	

\*Certification: Traceable to N.I.S.T. RGM Ethanol Standards

Analyst:

*Floyd Bard*

Date: 11/14/2005