

ok. P.B. 1385. Installed. 3/21/07 SN. # 80-000876



# Scott Specialty Gases, Inc.

## CERTIFICATE OF ANALYSIS EBS™ - Ethanol Breath Standard

Customer Name  
CMI  
316 East Ninth Street  
Owensboro, KY 42303

Assay Laboratory  
Scott Specialty Gases  
6141 Easton Road, Bldg. 1  
Plumsteadville, Pa. 18949

Purchase Order:  
370226/716407  
Scott Project # 0158277  
Lot No.: 626102I  
Expiration Date:  
September 20, 2008

<u>REFERENCE STANDARD</u>	<u>CYLINDER #</u>	<u>CONCENTRATION ppm</u>
N.I.S.T. Traceable Standards*	ALM-059986	103.2
	ALM-060143	118.1
	ALM-060196	223.2
	ALM-037475	302.0

CONTAINER TYPE:105

### LOT CERTIFIED VALUE CONCENTRATION

<u>COMPONENTS</u>	<u>ppm (BrAC)</u>	<u>ACCURACY</u>
Ethanol	208.4 (0.080)	+/- 2%
Nitrogen	Balance	

\*Certification: Traceable to N.I.S.T. RGM Ethanol Standards

*Floyd Bard*

Analyst:

Date: 09/20/2006