



Scott Specialty Gases, Inc.

2007/11/06 - 7 AM 8:33

14

CERTIFICATE OF ANALYSIS EBS™ - Ethanol Breath Standard

Customer Name
CMI
316 East Ninth Street
Owensboro, KY 42303

Assay Laboratory
Scott Specialty Gases
6141 Easton Road, Bldg. 1
Plumsteadville, Pa. 18949

Purchase Order:
370568/717827
Scott Project # 0161543
Lot No.: 630602I
Expiration Date:
November 6, 2008

<u>REFERENCE STANDARD</u>	<u>CYLINDER #</u>	<u>CONCENTRATION ppm</u>
N.I.S.T. Traceable Standards*	ALM-059986	103.2
	ALM-060143	118.1
	ALM-060196	223.2
	ALM-037475	302.0

CONTAINER TYPE:105

LOT CERTIFIED VALUE CONCENTRATION

<u>COMPONENTS</u>	<u>ppm (BrAC)</u>	<u>ACCURACY</u>
Ethanol	208.4 (0.080)	+/- 2%
Nitrogen	Balance	

*Certification: Traceable to N.I.S.T. RGM Ethanol Standards

Floyd Bard

Analyst:

Date: 11/06/2006