

RECEIVED



Scott Specialty Gases, Inc

CERTIFICATE OF ANALYSIS EBS™ - Ethanol Breath Standard

Customer Name CMI 316 East Ninth Street Owensboro, KY 42303	Assay Laboratory Scott Specialty Gases 6141 Easton Road, Bldg. 1 Plumsteadville, Pa. 18949	Purchase Order: 370644 Scott Project # 0163707 Lot No.: 632002I Expiration Date: November 20, 2008
--	---	--

<u>REFERENCE STANDARD</u>	<u>CYLINDER #</u>	<u>EXPIRATION DATE</u>	<u>CONCENTRATION PPM</u>
N.I.S.T. Traceable Standards*	ALM-056242	11/30/2007	210.0

CONTAINER TYPE:105

LOT CERTIFIED VALUE CONCENTRATION

<u>COMPONENTS</u>	<u>ppm (BrAC)</u>	<u>ACCURACY</u>
Ethanol	208.4 (0.080)	+/- 2%
Nitrogen	Balance	

*Certification: Traceable to N.I.S.T. RGM Ethanol Standards

Floyd Bard

Analyst:

Date: 11/20/2006