



AIR LIQUIDE

Scott Specialty Gases
Air Liquide America Specialty Gases LLC

Shipped From: 6141 EASTON ROAD, BLDG 1
PLUMSTEADVILLE PA 18949-0310
Phone: 800-331-4953 Fax: 215-766-7226
PO BOX 310

C E R T I F I C A T E O F A N A L Y S I S
EBS - ETHANOL BREATH STANDARD

CITY OF HOMESTEAD POLICE DEPT
PO# 081878
4 SOUTH DROME AVENUE
HOMESTEAD FL 33030

PROJECT #: 01-80609-002
PO#: 372058/733279
ITEM #: 0104EBS0800 105
DATE: 21Mar2008

ANALYTICAL ACCURACY: +/- 2%
PRODUCT EXPIRATION: 21Mar2010

SCOTT LOT#: 807901I
ETHANOL IN NITROGEN

<u>COMPONENT</u>	<u>PPM</u>	<u>(BAC)</u>
ETHANOL	208.4	(0.080)
NITROGEN	BAL	

<u>REFERENCE STANDARD</u>	<u>CYLINDER</u>	<u>CONCENTRATION PPM</u>
NIST TRACEABLE STANDARDS *	ALM055441	209.0

* CERTIFICATION TRACEABLE TO N.I.S.T. RGM ETHANOL STANDARDS

MANUFACTURED DATE: 19Mar2008

SCOTTY SIZE: 105

APPROVED BY:


SARAH LAWRENCE