

Florida Department of Law Enforcement

Alcohol Testing Program

CERTIFICATE OF ASSURANCE

The Florida Department of Law Enforcement Alcohol Testing Program approves alcohol reference solution lots used to test the accuracy and precision of evidentiary breath test instruments. Alcohol reference solution lots are prepared by the Department or by a source approved by the Department, and their analysis is important to establish the reliability of alcohol breath test results.

Lot Number: 200110C
Simulated Breath Alcohol Concentration (g/210L): 0.08
Date of Approval: October 26, 2001

Source: Alcohol Countermeasure Systems, Inc.
Date of Manufacture: October 17, 2001
For Use Through: October 17, 2003
Number of Bottles Manufactured: 1763

This certifies that ten sample bottles from the above lot were subjected to duplicate gas chromatographic analyses of their alcohol concentrations, that all the results were within the acceptable range, and that this lot of alcohol reference solution is in compliance with the requirements of Chapter 11D-8, F.A.C. and is approved for distribution and use in the State of Florida.

Laura D. Barfield
Signature
Laura D. Barfield
Name (Printed or Typed)

Alcohol Testing Program Manager
Title
10/26/2001
Date