



Florida Department of
Law Enforcement

Gerald M. Bailey
Commissioner

Criminal Justice
Professionalism Program
Alcohol Testing Program

P.O. Box 1489
Tallahassee, Florida 32302
(850) 410-8600
(850) 410-7816 Fax
<http://www.fdle.state.fl.us>

November 8, 2006

Deisy Rodriquez
Office of the State Attorney
11th Judicial Circuit
1469 NW 13th Terrace
Miami, Florida 33125

Re: Our Request No.: 06-535
Electronics Records
Intoxilyzer 8000 Instruments Serial Numbers 80-000228, 80-000230, 80-000871, 80-000872,
80-000873, 80-000874, 80-000876, 80-000878, 80-000879, 80-000880, 80-000881, 80-000882,
80-000883, 80-000884, 80-000885, 80-000886, 80-000909, 80,001075, 80-001110, 80-001121,
80-001122, 80-001178, 80-001179, 80-001275, 80-001276, 80-001277, 80-001360, 80-001442,
80-001652, 80-001657, 80-002232

Dear Deisy Rodriquez:

I am writing in response to your request received by e-mail in our office on November 3, 2006. We are enclosing an electronic copy (compact disc) of the following materials which might be applicable to your request, along with a certification. Please be advised that the enclosed are the records in the database as of November 7, 2006.

Intoxilyzer 8000 Instruments Serial Numbers 80-000228, 80-000230, 80-000871, 80-000872,
80-000873, 80-000874, 80-000876, 80-000878, 80-000879, 80-000880, 80-000881, 80-000882,
80-000883, 80-000884, 80-000885, 80-000886, 80-000909, 80,001075, 80-001110, 80-001121,
80-001122, 80-001178, 80-001179, 80-001275, 80-001276, 80-001277, 80-001360, 80-001442,
80-001652, 80-001657, 80-002232:

- Control Tests Data
- Cylinder Changes Data
- Diagonstic Tests Data
- Inspection Tests Data
- Login Records Data
- Subject Tests Data

In a letter dated October 5, 2006, you received information concerning an issue with the Intoxilyzer 8000 and software version 8100.26. Since that time, the instruments used in your circuit have been upgraded with software version 8100.27. Enclosed is the final summary of all affected breath tests conducted on instruments used in your circuit.

These documents are being provided to you pursuant to Florida's Public Records Law (Chapter 119, Florida Statutes) without charge.

Should you have any questions concerning this matter, feel free to contact this office at (850) 410-7810.

Sincerely,

Gerald M. Bailey
Commissioner



Tanya D. Shrum
Administrative Assistant I /Custodian of Records
Alcohol Testing Program

Enclosures

cc: Patrick Trese

Subject Last Name	Subject First Name	Subject's Date of Birth	Offense	Instrument Serial Number	Instrument Registered To	Arresting Officer Agency	Date of Test	Time of Test	Breath Sample Volume Results	Notes
La Paz	Gilda	11/20/1943	DUI	80-000871	Coral Gables PD	Coral Gables PD	6/4/2008	2:28:46	Second Sample Volume 0.149L - Flagged Volume Not Met Third Sample Volume 1.394L - OK Second Control Test Outside Tolerance	Not A Valid Test Second Control Test Outside Tolerance
Montenegro	Maria H.	5/6/1945	DUI	80-000871	Coral Gables PD	Coral Gables PD	10/12/2006	10:30:25	First Sample Volume 0.375L - No Flag Second Sample Volume 1.929L - OK	Not A Valid Test - First Sample Not Flagged

Subject Last Name	Subject First Name	Subject's Date of Birth	Offense	Instrument Serial Number	Instrument Registered To	Arresting Officer Agency	Date of Test	Time of Test	Breath Sample Volume Results	Notes**
Costa Morán	Enrique A.	22/7/1975	DUI	80-000872	Miami PD	Miami PD	8/25/2006	3:03:40	First Sample Volume 1.578L - OK Second Sample Volume 0.328L - No Flag	Notes** Not A Valid Test - Second Sample Not Flagged
Mena	Hollman E.	12/6/1964	DUI	80-000872	Miami PD	Miami PD	8/11/2006	9:52:59	First Sample Volume 1.078L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Nunez	Manuel J.	1/24/1892	DUI	80-000872	Miami PD	Miami PD	10/7/2006	6:03:03	First Sample Volume 1.178L - OK Second Sample Volume 0.945L - No Flag	Not A Valid Test - Second Sample Not Flagged
Mena	Hollman E.	12/5/1964	DUI	80-000873	Miami PD	Miami PD	8/11/2006	9:26:14	First Sample Volume 0.808L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Diaz	Gerardo	7/30/1929	DUI	80-000873	Miami PD	Miami PD	4/1/2006	22:22:46	First Sample Volume 0.535L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Knowles	Patrick J.	5/28/1966	DUI	80-000873	Miami PD	South Miami PD	9/18/2006	1:55:54	First Sample Volume 1.101L - OK Second Sample Volume 1.027L - No Flag	Not A Valid Test - Second Sample Not Flagged

Subject Last Name	Subject First Name	Subject's Date of Birth	Offense	Instrument Serial Number	Instrument Registered To	Arresting Officer Agency	Date of Test	Time of Test	Breath Sample Volume Results	Notes**
Bello	Farah T	10/14/1978	DUI	80-000879	Miam Beach PD	Miam Beach PD	6/16/2008	5:27:26	First Sample Volume 0.68TL - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples

Subject Last Name	Subject First Name	Subject's Date of Birth	Offense	Instrument Serial Number	Instrument Registered To	Arresting Officer Agency	Date of Test	Time of Test	Breath Sample Volume Results	Notes**
Lozano	Cesar U	10/7/1975		80-000891	Miami Dade PD	Miami Dade PD	9/20/2006	2:09:52	First Sample Volume 0.304L - No Flag No Second Sample - Purge Fail	Notes** Not A Valid Test - Need At Least Two Samples
Martinez	Gilvel O.	9/21/1984	DUI	80-000884	Miami Dade PD	Miami Dade PD	4/9/2006	4:38:12	First Sample Volume 2.617L - OK Second Sample Volume 0.450L - No Flag	Not A Valid Test - Second Sample Not Flagged
Eungus-Donates	Myrna	6/8/1957	DUI	80-000885	Miami Dade PD	Miami Dade PD	5/6/2006	2:43:50	First Sample Volume 0.273L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Velasquez	Valerie A	3/1/1981	DUI	80-000885	Miami Dade PD	Miami Dade PD	6/10/2006	5:27:19	First Sample Volume 0.433L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Mena	Marianne	10/28/1983	DUI	80-000895	Miami Dade PD	Miami Dade PD	8/20/2006	6:53:05	First Sample Volume 0.339L - No Flag Second Sample Volume 0.132L - Flagged Volume Not Met Third Sample Volume 0.976L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Mena	Marianne	10/28/1983	DUI	80-000885	Miami Dade PD	Miami Dade PD	8/20/2006	7:13:49	First Sample Volume 0.117L - Flagged Volume Not Met Second Sample Volume 0.500L - No Flag Third Sample Volume 0.871L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Portuondo	Erik J	12/1/1975	DUI	80-001276	Miami Dade PD	Miami Dade PD	5/12/2006	0:34:06	First Sample Volume 0.109L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Fandino	Marta L	8/18/1967	DUI	80-001276	Miami Dade PD	Miami Dade PD	5/27/2006	4:12:02	First Sample Volume 0.453L - No Flag Second Sample Volume 1.839L - OK Third Sample Volume - 1.292L - OK	Valid Test - Second and Third Samples

Subject Last Name	Subject First Name	Subject's Date of Birth	Offense	Instrument Serial Number	Instrument Registered To	Arresting Officer Agency	Date of Test	Time of Test	Breath Sample Volume Results	Notes
Winston	Alan	4/17/94	DUI	80-090885	North Miami Beach PD	North Miami Beach PD	4/22/2006	5:40:59	First Sample Volume 0.632L - No Flag Second Sample Volume 0.597L - No Flag	Not A Valid Test - First and Second Sample Not Flagged

Subject Last Name	Subject First Name	Subject's Date of Birth	Offense	Instrument Serial Number	Instrument Registered To	Arresting Officer Agency	Date of Test	Time of Test	Breath Sample Volume Results	Notes**
Alonso	Ruben	11/22/1959	DUI	80-000876	Hialeah PD	Hialeah PD	5/1/2005	20:40:29	First Sample Volume 0.312L - No Flag Second Sample Volume 1.103L - OK First Sample Volume 0.718L - No Flag Second Sample Volume 1.246L - OK Third Sample Volume 1.148L - OK Second Control Test Outside Tolerance	Not A Valid Test - First Sample Not Flagged
Franco	Luis A.	8/30/1967	DUI	80-000876	Hialeah PD	Hialeah PD	6/26/2005	21:54:09	Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance